



FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: EX650

Description: Clamp Meter, AC, 600A, True RMS + NCV

Date of Issue: 22-Jun-16

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

EMC Directive: 2014/30/EU
Report Number: R10983717-E1
Report Date of Issue: 11/6/2015

Standards:

IEC 61326-1:2012 (Second Edition)
IEC 61326-2-2:2012

LVD Directive 2014/35/EU
Report Number: E201687-A19-CB-1
Report Date of Issue: 3/7/2016

Standards:

IEC 61010-1:2010 (3rd Edition)

RoHS Directive: 2011/65/EU

Standard:

EN 50581:2012

The test reports show that the product fulfills the requirement in the EC Low Voltage Directive, EMC Directive, and RoHS Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC Low Voltage Directive, EMC Directive, and RoHS Directive.



Tony Campagna
Director of Quality Services