



FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: MM750W
Description: Multimeter w/BT
Date of Issue: 24-Apr-19

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

EMC Directive: 2014/30/EU
Report Number: R12252483-EMC
Report Date of Issue: 5/7/2018

Standards:
EN 61326-1:2013
EN 61326-2-2:2013

LVD Directive: 2014/35/EU
Report Number: E201687-D1014-1/A0/ULCB
Report Date of Issue: 11/12/2018

Standards:
IEC 61010-1:2010
IEC 61010-2-033:2012

R&TTE Directive: 2014/53/EU
Report Number: R122522483-EN 301 489
Report Date of Issue: 5/7/2018

Standards:
Draft EN 301 489-1 v2.2.0:2017
Draft EN 301 489-17 v3.2.0:2017

RoHS Directive: 2011/65/EU

Standard:
EN 50581:2012

The test reports show that the product fulfills the requirement in the EC Low Voltage Directive, EMC Directive, R&TTE Directive, and RoHS Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC Low Voltage Directive, EMC Directive, R&TTE Directive, and RoHS Directive.

Mark Sultzbach / QA Manager