



FLIR COMMERCIAL SYSTEMS, INC. – EXTECH BRAND
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: MO750
Description: Soil Moisture Meter
Date of Issue: 12-July-2012

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed below has been tested by a third party for CE-marking according to:

EMC Directive: 89/336/EEC
Report Number: 2008-080307-O.EMI
Certificate Date of Issue: March 7, 2008

Standards:
EN 61000-4-2
EN 61000-4-3

The test reports show that the product fulfills the requirement in the EC EMC Directive for CE marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC EMC Directive.


Mark Sultzbach (QA Administrator)
(for Tony Campagna)
Director of Quality Services