

THANK YOU

N99xxA/B FieldFox Handheld Analyzer Quick Start Guide

Congratulations, with this instrument you will achieve fast, accurate results with confidence.

CAUTION

IMPORTANT!

FULLY charge the FieldFox battery before using the FieldFox for the first time (i.e., The battery will be fully charged in about 4 hours).

When setting FieldFox on a surface, it is recommended to operate it with the kickstand open. The cooling is less effective when FieldFox is placed lying down flat on the instrument's back.

- Learn more or for questions: Download the N9927-90001 or N9938-90003 User's Guide.

- See also, www.keysight.com/find/fieldfox.

Step 1

1. Install the battery pack into the FieldFox battery compartment. The battery door is located on the lower-right side of the FieldFox (i.e., Press and slide the door down to open). For more, see Step 3.



Power connector



Audio jack

DC VoltageSource

Step 2

2. Insert the supplied AC/DC adapter (15-19 VDC) into the power connector and into an electrical outlet (AC/DC adapter is not shown).

- N9912A and N9923A models: The power connector is located on the right side, under the middle hinged door.

- All other models: The power connector is located on the LEFT side, under the bottom hinged cover.

Step 3

3. Extend the kickstand and press the power button.



Power button

LED

Kickstand

Battery door



keysight.com

Find a Manual · Register Your Instrument
Browse Application Notes
Request Repair or Calibration
View Discussion Forums · Updates



Keysight.com



This information is subject to change
without notice.

© Keysight Technologies 2014-2019

Edition 1, August 2019



N9927-90005

www.keysight.com